



ISO9001:2000 Registered
Since June 1995
Certification # 6756

Pressure Systems

Certificate of Compliance

This is to certify the described Software / Firmware has been developed, released, and tested in conformance with Pressure Systems, Inc. Quality Management System requirements.

Software / Firmware Description:

waterMonitor™ firmware version 2.3

Description of Change:

Firmware update from V 2.2 to 2.3. The reason for the change is to improve data volatility and reliability. This version permanently stores all measurements when taken. Previous versions waited until 512 bytes to approximately 35 measurements were taken before writing to permanent memory.


See verification document entitled "wateMONITOR Firmware Version 2.3 Data Volatility Improvement Verification Testing March 2008" All logging modes were tested and verified for proper operation and completion.

Notification must include normal website visibility, and customers should upgrade when possible.

PSI ECO Number:

#3951 Release of firmware version 2.3

Authorizing Signatures:

Design Engineering Manager:  Date: 7/25/08

Quality Assurance Manager:  Date: 7/25/08